



Ref. Certif. No.

DE 2-035853

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Genetic Sequencer
Name and address of the applicant	Illumina, Inc. 5200 Illumina Way San Diego CA 92122, USA
Name and address of the manufacturer	Illumina, Inc. 5200 Illumina Way San Diego CA 92122, USA
Name and address of the factory	Illumina Singapore Pte. Ltd. #02-13/18 29 Woodlands Industrial Park E1. North Tech Lobby Singapore 757716, Singapore
Ratings and principal characteristics	AC 100-240V, 50/60Hz; 240W; Class I; Laser Class 1
Trademark (if any)	illumina
Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	MiniSeq
Additional information (if necessary may also be reported on page 2)	Replaces CB Certificate US-TUVR-11250 due to change of NCB.
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010 IEC 61010-2-081:2015 IEC 60825-1:2014 for national differences see test report
As shown in the Test Report Ref. No. which forms part of this Certificate	31593844 015

This CB Test Certificate is issued by the National Certification Body



TÜV Rheinland LGA Products GmbH  
Tillystr. 2, 90431 Nürnberg, Germany  
Phone + 49 221 806-1371  
Fax + 49 221 806-3935  
Mail: cert-validity@de.tuv.com  
Web : www.tuv.com

Date: 2022-05-28

Signature:

Dipl.-Ing. M. Glagla