



Ref. Certif. No.

US-TUVR-009373-A3

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Genetic Sequencer
Name and address of the applicant	Illumina, Inc. 5200 Illumina Way San Diego CA 92122, USA
Name and address of the manufacturer	Illumina, Inc. 5200 Illumina Way San Diego CA 92122, USA
Name and address of the factory	Illumina Singapore Pte. Ltd. North Tech Lobby 3 #02-13/18 29 Woodlands Industrial Park E1 Singapore 757716, Singapore
Ratings and principal characteristics	AC 100-240V, 50/60Hz 240W; Class I Laser Class 1
Trademark (if any)	illumina
Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	MiniSeq
Additional information (if necessary may also be reported on page 2)	Re-issuance of CB Certificate US-TUVR-9373 due to 3rd modification (change of model/type ref., alternate components). See associated test report for details.
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010 IEC 61010-2-081:2015 IEC 60825-1:2014 for national deviations see test report
As shown in the Test Report Ref. No. which forms part of this Certificate	31593844 009

This CB Test Certificate is issued by the National Certification Body



TÜVRheinland®

TUV Rheinland of North America, Inc.
12 Commerce Road
Newtown, CT 06470 - USA

Date: 27.02.2018

Signature:

Martin Glagla